



FERRIS MFG. CORP.

5133 Northeast Parkway • Fort Worth, TX 76106-1822 • U.S.A.
Direct 817-900-1301 **1-800-POLYMEM** Fax 817-625-4277 info@polymem.com

July 3, 2025

PolyMem Silicone Border Dressing Modifications to IFU

To support the most common clinical applications and out of an abundance of caution related to potential ethylene oxide (EtO) sterilization residuals, Ferris Mfg. Corp. has updated the product labeling for PolyMem Silicone Border dressings.

- Since these dressings are most commonly used on adult patients, the following labeling changes have been made: **Product codes REF# 1248, 1249, 2439, 2466, 2468, and 2499** will now be labeled for use for **adults only, not for use for young children or neonates**.
- **Product codes REF# 1243, 1245, 2423, and 2435** will be labeled for use for **adults and children, but not for neonates**.

For use with pediatric or adult patients, always refer to the product's **Instructions for Use (IFU)** and consult your **sales representative** for additional guidance or information. The updated IFUs are in product shipments you received after June 25, 2025. The updated IFUs are also on the PolyMem website <https://polymem.com/instructions-for-use/>. If you require additional paper copies of updated IFUs to accompany products which you already hold, please email to orders@polymem.com and Ferris will provide these to you.

Thank you for using PolyMem dressings.

Paul Kay

Senior Vice President

www.polymem.com